

製造会社名〔 オムロン ヘルスケア株式会社 〕

記入日 2017年 8月 29日

	製品名	型番	試験結果に関する情報（学術論文等）	血圧計・カフのメンテナンス情報	掲載情報に関する連絡先
30	医用上腕式血圧計 (腕帯巻きつけタイプ)	HBP-T105S-N	<p>ESH International protocol</p> <p>● Roland Asmara, Jalil Belghazia, Jose ' Khabouthb, Jimmy Mattarb, Angelos Psimenosc and Giuseppe Germanoc. Validation of three professional devices measuring office blood pressure according to three different methods: the Omron BP10, the Omron HBP T105 and the Pic Indolor Professional. <i>Journal of Hypertension</i> March 2010 - Volume 28 - Issue 3 - p 452-458</p>	耐用期間 設置後6年	0120-103-203
31	医用上腕式血圧計 (腕帯巻きつけタイプ)	HEM-907	<p>ESH International protocol</p> <p>● Stefano Omboni, Isabella Riva, Alessia Giglio, Gianluca Caldara, Antonella Groppelli and Gianfranco Parati. Validation of the Omron M5-I, R5-I and HEM-907 automated blood pressure monitors in elderly individuals according to the International Protocol of the European Society of Hypertension. <i>Blood Pressure Monitoring</i> 2007, 12:233-242</p> <p>ESH International protocol</p> <p>● Mohamed A. El Assaad, Jirar A. Topouchian, Bernadette M. Darne and Roland G. Asmar. Validation of the Omron HEM-907 device for blood pressure measurement. <i>Blood Pressure Monitoring</i> 2002, 7:237-241</p> <p>AAMI</p> <p>● William B. White and Yusra Anis Anwar. Evaluation of the overall efficacy of the Omron office digital blood pressure HEM-907 monitor in adults. <i>Blood Pressure Monitoring</i> 2001, 6:107-110</p> <p>BHS</p> <p>● Gregory K. Davis, Lynne M. Roberts, Geroge J. Mangos and Mark A. Brown. Comparisons of auscultatory hybrid and automated sphygmomanometers with mercury sphygmomanometry in hypertensive and normotensive pregnant women: parallel validation studies. <i>Journal of Hypertension</i> 2015, 33: 499-506</p>	耐用期間 30,000回もしくは5年のいずれか早く到達した方	http://www.healthcare.omron.co.jp/support/

32	医用上腕式血圧計 (腕帯巻きつけタイプ)	HBP-1300	ANSI/AAMI/ISO81060-2:2013 ●Linghui Meng, Di Zhao, Yan Pan, Wenqing Ding, Qing Wei, Hua Li, Pingjin Gao and Jie Mi. Validation of Omron HBP-1300 professional blood pressure monitor based on auscultation in children and adults. <i>BMC Cardiovascular Disorders</i> 2016, 16:9	耐用期間 5年	0120-103-203
33	医用上腕式血圧計 (腕帯巻きつけタイプ)	HBP-1600	ANSI/AAMI/ISO81060-2:2013 ●Linghui Meng, Di Zhao, Yan Pan, Wenqing Ding, Qing Wei, Hua Li, Pingjin Gao and Jie Mi. Validation of Omron HBP-1300 professional blood pressure monitor based on auscultation in children and adults. <i>BMC Cardiovascular Disorders</i> 2016, 16:9	耐用期間 設置後6年	0120-103-203
34	生体情報モニタ (一体型)	BP-A308	ANSI/AAMI/ISO81060-2:2009 ●Mirna N. Chahine, PhD; Nathalie Assemani, MSc; Ghada Sayed Hassan, MD, Mariam Cham, MD, Pascale Salameh, PharmD, MPH, PhD; Roland Asmar, MD. Validation of the OMRON M3500 Blood Pressure Measuring Device Using Normal- and High-Speed Modes in Adult and Specific Populations (Obese and Children) According to AAMI Protocol. <i>The Journal of Clinical Hypertension</i> , Volume 17, Number 8, 1 August 2015, pp. 622-629(8)	耐用期間 設置後6年	0120-103-203
35	生体情報モニタ (一体型)	BP-A308D	ANSI/AAMI/ISO81060-2:2009 ●Mirna N. Chahine, PhD; Nathalie Assemani, MSc; Ghada Sayed Hassan, MD, Mariam Cham, MD, Pascale Salameh, PharmD, MPH, PhD; Roland Asmar, MD. Validation of the OMRON M3500 Blood Pressure Measuring Device Using Normal- and High-Speed Modes in Adult and Specific Populations (Obese and Children) According to AAMI Protocol. <i>The Journal of Clinical Hypertension</i> , Volume 17, Number 8, 1 August 2015, pp. 622-629(8)	耐用期間 設置後6年	0120-103-203
36	生体情報モニタ (一体型)	HBP-2070	ANSI/AAMI/ISO81060-2:2009 ●Mirna N. Chahine, PhD; Nathalie Assemani, MSc; Ghada Sayed Hassan, MD, Mariam Cham, MD, Pascale Salameh, PharmD, MPH, PhD; Roland Asmar, MD. Validation of the OMRON M3500 Blood Pressure Measuring Device Using Normal- and High-Speed Modes in Adult and Specific Populations (Obese and Children) According to AAMI Protocol. <i>The Journal of Clinical Hypertension</i> , Volume 17, Number 8, 1 August 2015, pp. 622-629(8)	耐用期間 設置後6年	0120-103-203

37	生体情報モニタ (一体型)	HBP-2071	ANSI/AAMI/ISO81060-2:2009 ●Mirna N. Chahine, PhD; Nathalie Assemani, MSc; Ghada Sayed Hassan, MD, Mariam Cham, MD, Pascale Salameh, PharmD, MPH, PhD; Roland Asmar, MD. Validation of the OMRON M3500 Blood Pressure Measuring Device Using Normal- and High- Speed Modes in Adult and Specific Populations (Obese and Children) According to AAMI Protocol. <i>The Journal of Clinical Hypertension</i> , Volume 17, Number 8, 1 August 2015, pp. 622-629(8)	耐用期間 設置後6年	0120-103-203
38	生体情報モニタ (一体型)	BP-608EV	ANSI/AAMI/ISO81060-2:2009 ●Mirna N. Chahine, PhD; Nathalie Assemani, MSc; Ghada Sayed Hassan, MD, Mariam Cham, MD, Pascale Salameh, PharmD, MPH, PhD; Roland Asmar, MD. Validation of the OMRON M3500 Blood Pressure Measuring Device Using Normal- and High- Speed Modes in Adult and Specific Populations (Obese and Children) According to AAMI Protocol. <i>The Journal of Clinical Hypertension</i> , Volume 17, Number 8, 1 August 2015, pp. 622-629(8)	耐用期間 設置後6年	0120-103-203